

VIASURE Viral SARS-CoV-2 Positive Control Kit

Ref. VS-VP1NCO

▶ **VIASURE Viral SARS-CoV-2 Positive Control Kit is intended for monitoring the whole process, from nucleic acid extraction to amplification for authorized WHO RT-qPCR protocols.**



▶ **Inactivated viral particles. Non infectious & Non replicative.**



▶ **Lyophilized presentation:** transport and store at room temperature.



▶ **Compatible with most common RNA extraction methods and WHO RT-qPCR protocols:** China CDC, US CDC (USA), National Institute of Infectious Diseases (Japan), Charité (Germany), HKU (Hong Kong SAR) and National Institute of Health (Thailand).



▶ **Monodose format exclusive for RT-qPCR assays.**

▶ **How does VIASURE Viral SARS-CoV-2 Positive Control Kit help your lab process?**

- Monitor instrument performance.
- Improve the diagnosis process: nucleic acid extraction, amplification and detection quality.
- Allows you to obtain comparable results.
- Validate and verify different assays complying with regulatory requirements.

▶ **Test procedure:**




**Reconstitution
of the
Positive Control**



**Nucleic acid
extraction**



Real-Time PCR



**Result
interpretation**

▶ **Reagents provided:**

Reagent/Material	Description	Amount
Viral SARS-CoV-2 Positive Control	Non-infectious lyophilized viral particle	4 vials
Viral Rehydration Buffer	Solution to reconstitute the lyophilized product	1 vial x 1 mL
Viral Negative Control	Non template control	4 vials x 1 mL

For more information and use procedure, read the instructions for use included in this product.

CerTest Biotec, S.L.

Pol. Industrial Río Gállego II · Calle J, Nº1
50840, San Mateo de Gállego, Zaragoza (Spain)
Tel. (+34) 976 520 354 · Fax (+34) 976 106 268
certest@certest.es | viasure@certest.es
www.certest.es



VIASURE/VP1NCO-0221EN